

Mesa Laboratories, Inc.  
12100 West 6th Ave., Lakewood, CO 80228

# Certificate of Conformance

Customer: Mesa Laboratories, Inc.  
Item: Thunder Scientific 2500 Humidity Generator  
ID Number: TE10228  
Serial Number: 9709106  
Procedure: Thunder Scientific User Manual Section 4  
Calibration Date: July 18, 2016  
Cert. Number: 1022818JUL2016



This certifies that the above product was calibrated in compliance with ISO/IEC 17025 using applicable Mesa Laboratories procedures. This calibration is traceable to the National Institute of Standards and Technology.

On the date tested, the instrument met its published operating specifications.

The environment in which this instrument was calibrated is maintained within the operating specification of the instrument and the standards.

U (Measurement Uncertainty) is the RSS (root sum square) of the standard deviation of the UUT's errors over the test range, the UUT's resolution and the uncertainty of the standard used, with a coverage factor of  $k=2$ .

A handwritten signature in blue ink, appearing to read 'H. Smith', written over a horizontal line.

Lab Manager

A handwritten signature in blue ink, appearing to read 'Jamie Jones', written over a horizontal line.

Quality Manager

This report shall not be reproduced except in full, without the written approval of Mesa Laboratories, Inc. Results relate only to the items tested or calibrated.

## REPORT OF CALIBRATION

Item: Thunder Scientific 2500 Humidity Generator S/N 9709106 , ID# TE10228  
 Low Range Pressure Transducer, S/N 565342  
 High Range Pressure Transducer, S/N 564872

Low Pressure Calibration Required: As Received @ ambient, 30 & 50 psiA,  
 As Left @ ambient, 20, 30, 40 & 50 psiA.

High Pressure Calibration Required: As Received @ 50, 100 & 150 psiA,  
 As Left @ 50, 75, 100, 125 & 150 psiA.

Calibration Date: 21-June-2016 @ 10 :00 AM/PM  
 Ambient Conditions: 21.18 °C and 33.39 %RH  
 Test Medium: Air

Standards Used: TE 10572 calibration expires: 12-January -2017  
 ±0.025% of reading uncertainty

### As Received Data:

Low Pressure Transducer				High Pressure Transducer			
Test psiA	2500 psiA	Error psiA	<b>U</b> ± psiA	Test psiA	2500 psiA	Error psiA	<b>U</b> ± psiA
12.00	11.97	0.03	0.01	50.00	50.01	0.01	0.11
30.00	29.98	0.02	0.01	100.00	100.00	0.00	0.11
50.00	49.98	0.02	0.02	146.00	146.00	0.00	0.12

### As Left Data:

Low Pressure Transducer				High Pressure Transducer			
Test psiA	2500 psiA	Error psiA	<b>U</b> ± psiA	Test psiA	2500 psiA	Error psiA	<b>U</b> ± psiA
12.00	12.00	0.00	0.01	50.00	49.99	0.01	0.02
20.00	19.99	0.01	0.01	75.00	74.99	0.01	0.02
30.00	30.00	0.00	0.01	100.00	99.99	0.01	0.03
40.00	40.00	0.00	0.01	125.00	124.99	0.01	0.04
50.00	50.00	0.00	0.02	146.00	146.00	0.00	0.04

Adjustments: New calibration coefficients were calculated and saved to memory


Manufacturer's specifications: ± 0.15% of full scale

As Received: Within Tolerance: YES  
 Operational Failure: NONE  
 Physical Damage: NONE

As Left: Within Tolerance: YES  
 Limited Range: NONE

This report shall not be reproduced except in full, without the written approval of Mesa Laboratories, Inc. Results relate only to the items tested or calibrated.

  
 Calibration Technician

  
 Lab Manager

## REPORT OF CALIBRATION

Item: Thunder Scientific 2500 Humidity Generator S/N 9709106, ID# TE 10228

Temperature Calibration Required: As Received @ 0, 35 & 70 °C  
As Left @ 0,20,35,50 & 70 °C

Calibration Date: 8-June-2016 @ 9:22 AM/PM  
Ambient Conditions: 23.56 °C and 30.50 %RH  
Test Medium: Dimethyl Silicone Fluid

Standards Used: Hart TE 10344                      calibration expires: 7/30/2016  
Hart TE    calibration expires:  
±0.018 °C uncertainty

**As Received Data:**

Actual °C	Saturation °C	Error °C	Exp Valve °C	Error °C	Chamber °C	Error °C	U ± °C
70.314	70.340	0.026	70.350	0.036	70.340	0.026	0.04
35.335	35.340	0.005	35.340	0.005	35.340	0.005	0.04
0.317	0.3450	0.028	0.3360	0.019	0.3650	0.048	0.04

**As Left Data:**

Actual °C	Saturation °C	Error °C	Exp Valve °C	Error °C	Chamber °C	Error °C	U ± °C
70.328	70.32	0.0080	70.32	0.0080	70.32	0.0080	0.03
60.336	60.33	0.0060	60.33	0.0060	60.33	0.0060	0.03
50.332	50.33	0.0020	50.33	0.0020	50.33	0.0020	0.03
40.335	40.35	0.0150	40.35	0.0150	40.35	0.0150	0.03
30.333	30.32	0.0130	30.32	0.0130	30.32	0.0130	0.03
20.328	20.32	0.0080	20.32	0.0080	20.32	0.0080	0.03
10.323	10.330	0.0070	10.330	0.0070	10.330	0.0070	0.03
0.317	0.3098	0.0072	0.3101	0.0069	0.3530	0.0360	0.03

Adjustments: New calibration coefficients were calculated and saved to memory  
Note: As per manufacturer's instructions the Presat temperature was not calibrated.


Manufacturer's specifications: ± 0.06 °C

As Received: Within Tolerance: YES  
Operational Failure: NONE  
Physical Damage: NONE

As Left: Within Tolerance: YES  
Limited Range: NONE

This report shall not be reproduced except in full, without the written approval of Mesa Laboratories, Inc. Results relate only to the items tested or calibrated.

  
\_\_\_\_\_  
Calibration Technician

  
\_\_\_\_\_  
Lab Manager

REPORT OF CALIBRATION

Item: Thermo Scientific 3500 Humidity Generator Part 6225-00 IOW 15 10000

Temperature Calibration: As Reported: 0.1, 35 & 70 °C  
As Test: 0.1, 35, 50 & 70 °C

Calibration Date: 5 June 2015 @ 0:52 AM EDT  
Amplified Location: 43.56 °C (see 3014) (WFI)  
Test Medium: Distilled Water

Standard Used: NIST 1013C  
Reference: NIST  
Calibration Error: 0.001296 °C  
Correction Error: 0.001296 °C

Setpoints°  
C

Setpoint (°C)	As Reported (°C)	Calibration Error (°C)	Correction Error (°C)
70	0.000676	0.001296	0.000676
30	2.5E-05	2.5E-05	2.5E-05
0	0.000784	0.000361	0.002304
	0.022249	0.023678	0.031649

Setpoints°  
C

Setpoint (°C)	As Reported (°C)	Calibration Error (°C)	Correction Error (°C)
70	6.4E-05	6.4E-05	6.4E-05
60	3.6E-05	3.6E-05	3.6E-05
50	4E-06	4E-06	4E-06
40	0.000225	0.000225	0.000225
30	0.000169	0.000169	0.000169
20	6.4E-05	6.4E-05	6.4E-05
10.000	4.9E-05	4.9E-05	4.9E-05
0.0000	5.18E-05	4.76E-05	0.001296
	0.009102	0.009073	0.015439