



This declaration of conformity is issued under the sole responsibility of the manufacturer.

IBP Medical GmbH
Ikarusallee 15
30179 Hannover, Germany

Object of the declaration:

| | | |
|-----------|-------------|---|
| Products: | HDU-CDTP | Conductivity/Temperature Sensor |
| | HDU-PRS | Pressure Sensor Family |
| | HDU-FL | Flow Sensor Family |
| | HDU-Pt100-I | Interface for Pt100 Temperature Sensors |
| | HDU-pH-I | Interface for pH-Sensor |
| | HDU-pHPt-I | Interface for pH-Sensor and Pt100 Temperature Sensors |

We hereby declare that the products mentioned above conform to the following European Directives:

Relevant EU Directives

EMC 2014/30/EU
LVD 2014/35/EU
RoHS 2015/863

Compliance with

EN 55011:2018-05
EN 61000-4-2:2009-12
EN 61000-4-3:2003-11
EN 61000-4-20:2011-07
EN 61326-1:2013-07
EN IEC 63000:2019-05

Hannover, March 04, 2025

Signed on behalf of IBP Medical GmbH, 15 Ikarusallee, Hannover 30179, Germany

Andrea Ladig
(Managing representative)