



# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## Certificate of Accreditation

*Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:*

***Mesa Laboratories, Inc.***  
12100 West 6<sup>th</sup> Avenue, Lakewood, CO 80228

*(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:*

**ISO/IEC 17025:2005**

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated January 2009):

***Chemical Testing for pH and Conductivity***  
*(As detailed in the supplement)*

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

*Initial Accreditation Date:*

January 7, 2010

*Issue Date:*

February 14, 2014

*Expiration Date:*

May 31, 2016

Tracy Szerszen  
President/Operations Manager

*Accreditation No.:*

66239

*Certificate No.:*

L14-46

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
755 W. Big Beaver, Suite 1325  
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: [www.pjlab.com](http://www.pjlab.com)*



# Certificate of Accreditation: Supplement

## Mesa Laboratories, Inc.

12100 West 6<sup>th</sup> Avenue, Lakewood, CO 80228  
Jamie Louie Phone: 303-987-8000

*Accreditation is granted to the facility to perform the following tests:*

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Chemical	pH Calibrator Solutions	pH	EPA Method 150.1	pH 4.00 to pH 10.00 DL = 0.001 pH
	Conductivity Calibrator Solutions	Conductivity	EPA Method 120.1	0.150 mS/cm to 100 mS/cm DL = 0.0001 @ 0.150 mS/cm DL = 0.0002 @ 1.000 mS/cm DL = 0.0016 @ 13.5 to 14.0 mS/cm DL = 0.0335 @ 100.0 mS/cm

