

PRODUCT CHANGE NOTIFICATION

27 November 2018

Notification ID: CN-181003


Notification Type: Product Labeling Change

Dear Customer,

We would like to take this opportunity to notify you of a change that will be made to the Indicator Card for **SRSP/25 Smart-Read Steam Log 5 Test Pack 25/25C**. The following changes will be made to the indicator card:

- The EZTest and MesaLabs logos will be removed.
- The process will be added (steam).
- The results to be scored by the user will be “Growth” or “No Growth” as opposed to “Achieved” or “Not Achieved”.

The changes are being made for business reasons. They are considered cosmetic, as they do not impact the use of the product and there will be no change to the function of the card. Examples of the current card and new card are provided below.

Current Indicator Card	New Indicator Card
<p>EZTest® INSTANT READOUT INTEGRATOR</p> <p>STERILIZER NO.: _____ LOAD NO.: _____</p> <p>DATE: _____</p> <div style="border: 1px solid black; padding: 5px; display: inline-block;"> <p>PASS 270 °F</p> <p>IF GREEN COLOR STANDARD 000000</p> </div> <p>RESULTS: STERILIZATION CONDITIONS</p> <p><input type="checkbox"/> ACHIEVED <input type="checkbox"/> NOT ACHIEVED</p> <p>DATE: _____</p> <p>RELEASED BY: _____</p>  <p>MesaLabs</p>	<p>INSTANT READOUT INTEGRATOR</p> <p>STERILIZER NO _____ LOAD NO _____ DATE _____</p> <div style="border: 1px solid black; padding: 5px; display: inline-block;"> <p>PASS 270 °F</p> <p>IF GREEN COLOR STANDARD 000000</p> </div> <p>STEAM</p> <p>CONVENTIONAL BI RESULTS</p> <p><input type="radio"/> GROWTH - REJECT</p> <p><input type="radio"/> NO GROWTH - ACCEPT</p> <p>DATE _____</p> <p>RELEASED BY _____</p>

Mesa will continue to distribute **SRSP/25** with the current Indicator Card until current stock is consumed. At current rates of distribution, we expect the change to occur after 31 December 2018. As always, if you have any questions, please contact your Mesa Laboratories Representative.

 _____ 27 Nov 2018
 Robert Bradley, Director – Manufacturing Date

 _____ 27 Nov 2018
 Janis E. Smoke, Director – Quality & Regulatory Affairs Date